

Ref	Item	Rating	Comments
1 (Section 4.1)	Has an EPA been appointed?		
2 (Section 4.2.1)	Has the EPA provided advice to management?		
3 (Section 4.2.3)	Has the EPA conducted surveys of static magnetic fields?		
4 (Sections 4.3.1 and 4.4.1)	Have exposure assessments been made?		
5 (Sections 4.3.3 and 4.4.3)	Have accessible fields been considered for equipment constructed in house?		
6 (Sections 4.3.4 and 4.4.4)	Are risk assessments documented for areas where work is undertaken in strong static magnetic fields?		

7 (Section 4.4.4)	Has a risk assessment been made and plans drawn up for assembly of equipment containing strong permanent magnets?		
8 (Section 4.3.5)	Have warning signs been posted at entrances to areas where field strengths exceed 0.5 mT?		
9 (Section 4.4.8)	Is the 3mT (30 Gauss) perimeter of strong static magnetic fields marked/bounded?		
10 (Section 4.4.8)	Are warning signs for the presence of strong static magnetic fields posted at the entrances to areas where strong static magnetic fields are present?		
11	Have local procedures for the safe operation of equipment been developed for work undertaken in areas where strong static magnetic fields exist?		